
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: G775
PDI Gage Factor: 94.3
Mean Linear Correlation Coefficient: 0.999916

Table 1: Representative Calibration Data

Applied strain (με)	Gage Output (mV/Vexc)	Applied strain (με)	Gage Output (mV/Vexc)
-1934.8	-5.896	525.1	1.700
-1507.6	-4.594	930.7	2.930
-1090.6	-3.317	1317.0	4.077
-640.7	-1.951	1677.5	5.164
-229.5	-0.678	1993.3	6.072
131.3	0.486		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1006HA
Firmware version number:	0.9.1.1
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Paul Hartman

Signature: Paul Hartman

PDI Gage: G775

Calibration Date: 10/26/2026 9:32:05 AM

Calibration valid 2 years

Pile Dynamics, Inc.